

EXHIBIT I-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
INTOXILYZER MODEL 9000

PERIODIC MAINTENANCE AND STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST Michael Guilbault AGENCY GOODYEAR POLICE DEPT
DATE 12/01/2024 TIME 1158
INTOXILYZER SERIAL # 90-004605

- ☒ 1. Ensure that gas tank is attached and contains a standard alcohol concentration .100 AC

DIAGNOSTIC TESTS

- ☒ 1. Clock time check
☒ 2. Date check


OPERATIONAL TESTS

- ☒ 1. Deficient Subject Test (Proper Sample Recognition)
Deficient Sample printed
- ☒ 2. Alcohol-free Subject Test (Proper Sample Recognition):
0. 000 AC
- ☐ 3. Mouth Alcohol Subject Test (Proper Sample Recognition): — see second page
- ☒ 4. Radio Frequency Interference Test (Error Recognition):
Invalid Sample – Begin new deprivation period printed
RFI Detect printed
- ☐ 5. Standard Calibration Check:
0. 102 AC
- ☒ 6. Air Blanks Completed
- ☒ 7. Timer Reset

Not a Successfully Completed Test Sequence will be printed

Instrument is operating properly and accurately. YES ☒ NO ☐

COMMENTS:

SIGNATURE  1144

ARIZONA FORENSIC BREATH ALCOHOL ANALYTICAL REPORT

INSTRUMENT INFORMATION

Analytical Instrument: Intoxilyzer 9000
Serial Number: 90-004605
Software Version: 9439.01.00
Analytical Report Number: 46051201240285

QAS: GUILBAULT, MICHAEL
QAS Permit #: 30305
Agency: GOODYEAR PD
Last 31-Day Check: 12/01/2024
Last Annual Maintenance: 09/20/2024

SUBJECT INFORMATION

Name: 31-Day Check
Test Date: 12/01/2024
Driver's License #:
Gender:
Date of Birth:
Age:

Weight:
State of Issue:
Driver's License Expiration:
Deprivation Start Time: 11:40
15 - Minute Deprivation: Yes

OPERATOR INFORMATION

Name: GUILBAULT, MICHAEL
Agency: GOODYEAR PD

Permit #: 30305
STANDARD INFORMATION
Standard Value: 0.100
Standard Lot #: 24923100A1
Expiration Date: 10/05/2025
Bottle #: 57
Last Changed By: MIYAZATO, KORY

Permit #: 30215

Test	g/210L	Time
Air Blank	0.000	11:58:26
Diagnostic Test	PASS	11:59:02
Air Blank	0.000	11:59:39
Calibration Chk	0.102	12:00:00
Air Blank	0.000	12:00:40
Subject Test 1	DEF*	12:04:03
Air Blank	0.000	12:04:46
Air Blank	0.000	12:05:26
Subject Test 2	0.000	12:05:50
Air Blank	0.000	12:06:32
Wait		12:10:35
Air Blank	0.000	12:11:12
Subject Test 3	0.029**	12:13:08
Air Blank	RFI***	12:13:36
Air Blank	0.000	12:14:13

*Deficient Sample
**No .020 Agreement
***RFI Detect

RESULTS

EXCEPTION MESSAGES

* Deficient Sample
** No .020 Agreement
*** RFI Detect

OPERATOR COMMENTS

Not a Successfully Completed Test Sequence

All air blank results must be 0.000.
Consecutive subject test results must not differ by more than 0.020 g/210L.
Standard check results must be $\pm 10\%$.

Intoxilyzer - Alcohol Analyzer
Model 9000

SN: 90-004605

Test Date: 12/01/2024

Location:

Test Time: 12:14

Software Version: 9439.01.00

Standard Lot #: 24923100A1

Standard Value: 0.100

Expiration Date: 10/05/2025

Bottle #: 57

Last Changed By: MIYAZATO, KORY

Permit #: 30215

Operator: GUILBAULT, MICHAEL

Permit #: 30305

Agency: GOODYEAR PD

Test	g/210L	Time
Air Blank	0.000	12:15:10
Calibration Chk	0.102	12:15:32
Air Blank	0.000	12:16:11

No RFI Detected

ARIZONA FORENSIC BREATH ALCOHOL ANALYTICAL REPORT

INSTRUMENT INFORMATION

Analytical Instrument: Intoxilyzer 9000
Serial Number: 90-004605
Software Version: 9439.01.00
Analytical Report Number: 46051201240286

QAS: GUILBAULT, MICHAEL
QAS Permit #: 30305
Agency: GOODYEAR PD
Last 31-Day Check: 12/01/2024
Last Annual Maintenance: 09/20/2024

SUBJECT INFORMATION

Name: 31-Day Check
Test Date: 12/01/2024
Driver's License #:
Gender:
Date of Birth:
Age:

Weight:
State of Issue:
Driver's License Expiration:
Deprivation Start Time: 12:30
15 - Minute Deprivation: Yes

OPERATOR INFORMATION

Name: GUILBAULT, MICHAEL
Agency: GOODYEAR PD

Permit #: 30305
STANDARD INFORMATION
Standard Value: 0.100
Standard Lot #: 24923100A1
Expiration Date: 10/05/2025
Bottle #: 57
Last Changed By: MIYAZATO, KORY
Permit #: 30215

Test	g/210L	Time
Air Blank	0.000	13:06:27
Diagnostic Test	PASS	13:07:03
Air Blank	0.000	13:07:39
Calibration Chk	0.102	13:08:01
Air Blank	0.000	13:08:40
Subject Test 1	INV*	13:08:57
Air Blank	0.000	13:09:54

*Invalid Sample - Begin new deprivation period

RESULTS

EXCEPTION MESSAGES

* Invalid Sample - Begin new deprivation period

OPERATOR COMMENTS

No RFI Detected

Not a Successfully Completed Test Sequence

*All air blank results must be 0.000.
Consecutive subject test results must not differ by more than 0.020 g/210L.
Standard check results must be $\pm 10\%$.*

Intoxilyzer - Alcohol Analyzer
Model 9000

SN: 90-004605

Test Date: 12/01/2024

Test Time: 13:10

Location:

Software Version: 9439.01.00

Standard Lot #: 24923100A1

Standard Value: 0.100

Expiration Date: 10/05/2025

Bottle #: 57

Last Changed By: MIYAZATO, KORY

Permit #: 30215

Operator: GUILBAULT, MICHAEL

Permit #: 30305

Agency: GOODYEAR PD

Test	g/210L	Time
Air Blank	0.000	13:11:07
Calibration Chk	0.101	13:11:29
Air Blank	0.000	13:12:09

No RFI Detected