

EXHIBIT I-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
INTOXILYZER MODEL 9000

PERIODIC MAINTENANCE AND STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST K Miyazawa 1139 AGENCY Goodyear
DATE 11/31/25 TIME 0029
INTOXILYZER SERIAL # 90-004605

☒ 1. Ensure that gas tank is attached and contains a standard alcohol concentration .100 AC

DIAGNOSTIC TESTS

- ☒ 1. Clock time check
☒ 2. Date check

OPERATIONAL TESTS

- ☒ 1. Deficient Subject Test (Proper Sample Recognition)
Deficient Sample printed
☒ 2. Alcohol-free Subject Test (Proper Sample Recognition):
0. 0.100 AC
☒ 3. Mouth Alcohol Subject Test (Proper Sample Recognition):
Invalid Sample – Begin new deprivation period printed
☒ 4. Radio Frequency Interference Test (Error Recognition):
RFI Detect printed
☒ 5. Standard Calibration Check:
0. 100 AC
☒ 6. Air Blanks Completed
☒ 7. Timer Reset

Not a Successfully Completed Test Sequence will be printed

Instrument is operating properly and accurately. YES ☒ NO ☐

COMMENTS:

SIGNATURE K Miyazawa 1139

ARIZONA FORENSIC BREATH ALCOHOL ANALYTICAL REPORT

INSTRUMENT INFORMATION

Analytical Instrument: Intoxilyzer 9000
Serial Number: 90-004605
Software Version: 9439.01.00
Analytical Report Number: 46050131250364

QAS: MIYAZATO, KORY
QAS Permit #: 30215
Agency: GOODYEAR PD
Last 31-Day Check: 12/31/2024
Last Annual Maintenance: 09/20/2024

SUBJECT INFORMATION

Name: 31-Day Check
Test Date: 01/31/2025
Driver's License #:
Gender:
Date of Birth:
Age:

Weight:
State of Issue:
Driver's License Expiration:
Deprivation Start Time: 00:10
15 - Minute Deprivation: Yes

OPERATOR INFORMATION

Name: MIYAZATO, KORY
Agency: GOODYEAR PD

Permit #: 30215

STANDARD INFORMATION

Standard Value: 0.100
Standard Lot #: 24923100A1
Expiration Date: 10/05/2025
Bottle #: 57
Last Changed By: MIYAZATO, KORY

Permit #: 30215

Test	g/210L	Time
Air Blank	0.000	00:29:39
Diagnostic Test	PASS	00:30:15
Air Blank	0.000	00:30:52
Calibration Chk	0.099	00:31:13
Air Blank	0.000	00:31:53
Subject Test 1	DEF*	00:35:04
Air Blank	0.000	00:35:44
Air Blank	0.000	00:36:26
Subject Test 2	0.000	00:36:43
Air Blank	0.000	00:37:23
Wait		00:41:26
Air Blank	0.000	00:42:03
Subject Test 3	INV**	00:42:25
Air Blank	RFI***	00:42:40

RESULTS

EXCEPTION MESSAGES

*Deficient Sample
**Invalid Sample - Begin new deprivation period
***RFI Detect

* Deficient Sample
** Invalid Sample - Begin new deprivation period
*** RFI Detect

OPERATOR COMMENTS

Not a Successfully Completed Test Sequence

*All air blank results must be 0.000.
Consecutive subject test results must not differ by more than 0.020 g/210L.
Standard check results must be $\pm 10\%$.*

Intoxilyzer - Alcohol Analyzer
Model 9000

SN: 90-004605

Test Date: 01/31/2025

Test Time: 00:43

Location:

Software Version: 9439.01.00

Standard Lot #: 24923100A1

Standard Value: 0.100

Expiration Date: 10/05/2025

Bottle #: 57

Last Changed By: MIYAZATO, KORY

Permit #: 30215

Operator: MIYAZATO, KORY

Permit #: 30215

Agency: GOODYEAR PD

Test	g/210L	Time
Air Blank	0.000	00:43:35
Calibration Chk	0.100	00:43:57
Air Blank	0.000	00:44:36

No RFI Detected

Intoxilyzer - Alcohol Analyzer
Model 9000

SN: 90-004605

Test Date: 02/26/2025

Test Time: 23:55

Location:

Software Version: 9439.01.00

Standard Lot #: 24923100A1

Standard Value: 0.100

Expiration Date: 10/05/2025

Bottle #: 54

Last Changed By: MIYAZATO, KORY

Permit #: 30215

Operator: MIYAZATO, KORY

Permit #: 30215

Agency: GOODYEAR PD

Test	g/210L	Time
Air Blank	0.000	23:55:24
Calibration Chk	0.102	23:55:45
Air Blank	0.000	23:56:25

No RFI Detected

Successfully Completed Test Sequence