

EXHIBIT I-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
INTOXILYZER MODEL 9000

PERIODIC MAINTENANCE AND STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST K Miyazato 1139 AGENCY Goodyear
DATE 3/4/25 TIME 072231
INTOXILYZER SERIAL # 90-03795

- ☒ 1. Ensure that gas tank is attached and contains a standard alcohol concentration -100 AC

DIAGNOSTIC TESTS

- ☒ 1. Clock time check
☒ 2. Date check

OPERATIONAL TESTS

- ☒ 1. Deficient Subject Test (Proper Sample Recognition)
Deficient Sample printed
☒ 2. Alcohol-free Subject Test (Proper Sample Recognition):
0. 000 AC
☒ 3. Mouth Alcohol Subject Test (Proper Sample Recognition):
Invalid Sample – Begin new deprivation period printed
☒ 4. Radio Frequency Interference Test (Error Recognition):
RFI Detect printed
☒ 5. Standard Calibration Check:
0. 099 AC
☒ 6. Air Blanks Completed
☒ 7. Timer Reset

Not a Successfully Completed Test Sequence will be printed

Instrument is operating properly and accurately. YES ☒ NO ☐

COMMENTS:

SIGNATURE Kunt 1139

ARIZONA FORENSIC BREATH ALCOHOL
ANALYTICAL REPORT
10701 S ESTRELLA PKY

Intoxilyzer - Alcohol Analyzer
Model 9000
SN: 90-003795
Test Date: 03/05/2025 Test Time: 22:44
Location: 10701 S ESTRELLA PKY

INSTRUMENT INFORMATION
Analytical Instrument: INTOXILYZER 9000
Serial Number: 90-003795
Software Version: 9439.01.00
Analytical Report Number: 37950305250717
QAS: GUILBAULT, MICHAEL
QAS Permit #: 30305
Agency: GOODYEAR PD
Last 31-Day Check: 12/05/2024
Last Annual Maintenance: 10/02/2024

Software Version: 9439.01.00
Standard Lot #: 302-402972207
Standard Value: 0.100
Expiration Date: 02/20/2026
Bottle #: 19
Last Changed By: MIYAZATO, KORY
Permit #: 30215

SUBJECT INFORMATION
Test Date: 03/05/2025
Name: 31-Day Check,
Driver's License #:
Weight:
Gender:
State of Issue:
Date of Birth:
Driver's License Expiration:
Age:
Deprivation Start Time: 22:00
15-Minute Deprivation: Yes

Operator: MIYAZATO, KORY
Permit #:30215
Agency: GOODYEAR PD

Test	g/210L	Time
Air Blank	0.000	22:45:06
Calibration Chk	0.099	22:45:27
Air Blank	0.000	22:46:07

No RFI Detected

OPERATOR INFORMATION
Name: MIYAZATO, KORY
Agency: GOODYEAR PD
Permit #: 30215

STANDARD INFORMATION
Standard Value: 0.100
Standard Lot #: 302-402972207
Expiration Date: 02/20/2026
Bottle #: 19
Last Changed By:
MIYAZATO, KORY
Permit #: 30215

Test	g/210L	Time
Air Blank	0.000	22:31:10
Diagnostic Test	PASS	22:31:45
Air Blank	0.000	22:32:22
Calibration Chk	0.098	22:32:44
Air Blank	0.000	22:33:25
Subject Test 1	DEF*	22:36:36
Air Blank	0.000	22:37:18
Air Blank	0.000	22:37:59
Subject Test 2	0.000	22:38:16
Air Blank	0.000	22:38:56
Wait		22:43:01
Air Blank	0.000	22:43:37
Subject Test 3	INV**	22:43:56
Air Blank	RFI***	22:44:14

*Deficient Sample
**Invalid Sample - Begin new deprivation period
***RFI Detect

Not a Successfully Completed Test Sequence

RESULTS

EXCEPTION MESSAGES

* Deficient Sample
** Invalid Sample - Begin new deprivation period
*** RFI Detect
All air blank results must be 0.000.
Consecutive subject test results must not differ
by more than 0.020 g/210L.
Standard check results must be ±10%.