

EXHIBIT I-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
INTOXILYZER MODEL 9000

PERIODIC MAINTENANCE AND STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST K Miyazato 1139 AGENCY Goodyear
DATE 3/29/25 TIME 0436
INTOXILYZER SERIAL # 90-005394

- ☒ 1. Ensure that gas tank is attached and contains a standard alcohol concentration .100 AC

DIAGNOSTIC TESTS

- ☒ 1. Clock time check
☒ 2. Date check

OPERATIONAL TESTS

- ☒ 1. Deficient Subject Test (Proper Sample Recognition)
Deficient Sample printed
☒ 2. Alcohol-free Subject Test (Proper Sample Recognition):
0. 000 AC
☒ 3. Mouth Alcohol Subject Test (Proper Sample Recognition):
Invalid Sample – Begin new deprivation period printed
☒ 4. Radio Frequency Interference Test (Error Recognition):
RFI Detect printed
☒ 5. Standard Calibration Check:
0. 100 AC
☒ 6. Air Blanks Completed
☒ 7. Timer Reset

Not a Successfully Completed Test Sequence will be printed

Instrument is operating properly and accurately. YES ☒ NO ☐

COMMENTS:

SIGNATURE Kms 1139

Intoxilyzer - Alcohol Analyzer

Model 9000

SN: 90-005394

Test Date: 03/29/2025

Test Time: 04:50

Location:

Software Version: 9439.01.00

Standard Lot #: 24923100A1

Standard Value: 0.100

Expiration Date: 10/05/2025

Bottle #: 51

Last Changed By: MIYAZATO, KORY

Permit #: 30215

Operator: MIYAZATO, KORY

Permit #: 30215

Agency: GOODYEAR PD

Test	g/210L	Time
Air Blank	0.000	04:50:24
Calibration Chk	0.100	04:50:46
Air Blank	0.000	04:51:27

No RFI Detected

ARIZONA FORENSIC BREATH ALCOHOL ANALYTICAL REPORT

INSTRUMENT INFORMATION

Analytical Instrument: Intoxilyzer 9000
Serial Number: 90-005394
Software Version: 9439.01.00
Analytical Report Number: 53940329250049

QAS: MIYAZATO, KORY
QAS Permit #: 30215
Agency: GOODYEAR PD
Last 31-Day Check: 03/01/2025
Last Annual Maintenance: 08/30/2024

SUBJECT INFORMATION

Name: 31-Day Check
Test Date: 03/29/2025
Driver's License #:
Gender:
Date of Birth:
Age:

Weight:
State of Issue:
Driver's License Expiration:
Deprivation Start Time: 04:00
15 - Minute Deprivation: Yes

OPERATOR INFORMATION

Name: MIYAZATO, KORY
Agency: GOODYEAR PD

Permit #: 30215

STANDARD INFORMATION

Standard Value: 0.100
Standard Lot #: 24923100A1
Expiration Date: 10/05/2025
Bottle #: 51
Last Changed By: MIYAZATO, KORY

Permit #: 30215

Test	g/210L	Time
Air Blank	0.000	04:36:19
Diagnostic Test	PASS	04:36:54
Air Blank	0.000	04:37:31
Calibration Chk	0.099	04:37:53
Air Blank	0.000	04:38:33
Subject Test 1	DEF*	04:41:46
Air Blank	0.000	04:42:31
Air Blank	0.000	04:43:11
Subject Test 2	0.000	04:43:28
Air Blank	0.000	04:44:08
Wait		04:48:12
Air Blank	0.000	04:48:48
Subject Test 3	INV**	04:49:09
Air Blank	RFI***	04:49:34

RESULTS

EXCEPTION MESSAGES

*Deficient Sample
**Invalid Sample - Begin new deprivation period
***RFI Detect

* Deficient Sample
** Invalid Sample - Begin new deprivation period
*** RFI Detect

OPERATOR COMMENTS

Not a Successfully Completed Test Sequence

All air blank results must be 0.000.
Consecutive subject test results must not differ by more than 0.020 g/210L.
Standard check results must be $\pm 10\%$.