

EXHIBIT I-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
INTOXILYZER MODEL 9000

PERIODIC MAINTENANCE AND STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST K Miyazato 1139 AGENCY GOODYEAR  
DATE 5/14/25 TIME 2223  
INTOXILYZER SERIAL # 90-003795

- ☒ 1. Ensure that gas tank is attached and contains a standard alcohol concentration 0.100 AC

DIAGNOSTIC TESTS

- ☒ 1. Clock time check  
☒ 2. Date check

OPERATIONAL TESTS

- ☒ 1. Deficient Subject Test (Proper Sample Recognition)  
Deficient Sample printed  
☒ 2. Alcohol-free Subject Test (Proper Sample Recognition):  
0.000 AC  
☒ 3. Mouth Alcohol Subject Test (Proper Sample Recognition):  
Invalid Sample – Begin new deprivation period printed  
☒ 4. Radio Frequency Interference Test (Error Recognition):  
RFI Detect printed  
☒ 5. Standard Calibration Check:  
0.096 AC  
☒ 6. Air Blanks Completed  
☒ 7. Timer Reset

Not a Successfully Completed Test Sequence will be printed

Instrument is operating properly and accurately. YES ☒ NO ☐

COMMENTS:

\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

SIGNATURE Kmt 1139

ARIZONA FORENSIC BREATH ALCOHOL  
ANALYTICAL REPORT

10701 S ESTRELLA PKY

INSTRUMENT INFORMATION

Analytical Instrument: INTOXILYZER 9000  
Serial Number: 90-003795  
Software Version: 9439.01.00  
Analytical Report Number: 37950514250719  
QAS: MIYAZATO, KORY  
QAS Permit #: 30215  
Agency: GOODYEAR PD  
Last 31-Day Check: 03/05/2025  
Last Annual Maintenance: 10/02/2024

SUBJECT INFORMATION

Test Date: 05/14/2025  
Name: 31-Day Check,  
Driver's License #:  
Weight:  
Gender:  
State of Issue:  
Date of Birth:  
Driver's License Expiration:  
Age:  
Deprivation Start Time: 22:00  
15-Minute Deprivation: Yes

OPERATOR INFORMATION

Name: MIYAZATO, KORY  
Agency: GOODYEAR PD  
Permit #: 30215

STANDARD INFORMATION

Standard Value: 0.100  
Standard Lot #: 302-402972207  
Expiration Date: 02/20/2026  
Bottle #: 19  
Last Changed By:  
MIYAZATO, KORY  
Permit #: 30215

| Test            | g/210L | Time     |
|-----------------|--------|----------|
| Air Blank       | 0.000  | 22:23:48 |
| Diagnostic Test | PASS   | 22:24:24 |
| Air Blank       | 0.000  | 22:25:01 |
| Calibration Chk | 0.097  | 22:25:25 |
| Air Blank       | 0.000  | 22:26:04 |
| Subject Test 1  | DEF*   | 22:29:20 |
| Air Blank       | 0.000  | 22:30:03 |
| Air Blank       | 0.000  | 22:30:43 |
| Subject Test 2  | 0.000  | 22:30:59 |
| Air Blank       | 0.000  | 22:31:40 |
| Wait            |        | 22:35:45 |
| Air Blank       | 0.000  | 22:36:22 |
| Subject Test 3  | INV**  | 22:36:42 |
| Air Blank       | RFI*** | 22:36:55 |

\*Deficient Sample  
\*\*Invalid Sample - Begin new deprivation period  
\*\*\*RFI Detect

Not a Successfully Completed Test Sequence

RESULTS

EXCEPTION MESSAGES

\* Deficient Sample  
\*\* Invalid Sample - Begin new deprivation period  
\*\*\* RFI Detect  
All air blank results must be 0.000.  
Consecutive subject test results must not differ  
by more than 0.020 g/210L.  
Standard check results must be ±10%.

Intoxilyzer - Alcohol Analyzer  
Model 9000

SN: 90-003795  
Test Date: 05/14/2025  
Test Time: 22:37

Location:  
10701 S ESTRELLA PKY

Software Version: 9439.01.00

Standard Lot #: 302-402972207  
Standard Value: 0.100  
Expiration Date: 02/20/2026  
Bottle #: 19  
Last Changed By: MIYAZATO, KORY  
Permit #: 30215

Operator: MIYAZATO, KORY  
Permit #:30215  
Agency: GOODYEAR PD

| Test            | g/210L | Time     |
|-----------------|--------|----------|
| Air Blank       | 0.000  | 22:37:53 |
| Calibration Chk | 0.096  | 22:38:14 |
| Air Blank       | 0.000  | 22:38:54 |

No RFI Detected