

EXHIBIT I-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
INTOXILYZER MODEL 9000

PERIODIC MAINTENANCE AND STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST K Miyazato 1139 AGENCY Goodyear
DATE 6/29/25 TIME 0353
INTOXILYZER SERIAL # 90-005394

☒ 1. Ensure that gas tank is attached and contains a standard alcohol concentration ~100 AC

DIAGNOSTIC TESTS

- ☒ 1. Clock time check
☒ 2. Date check

OPERATIONAL TESTS

- ☒ 1. Deficient Subject Test (Proper Sample Recognition)
Deficient Sample printed
- ☒ 2. Alcohol-free Subject Test (Proper Sample Recognition):
0. 000 AC
- ☒ 3. Mouth Alcohol Subject Test (Proper Sample Recognition):
Invalid Sample – Begin new deprivation period printed
- ☒ 4. Radio Frequency Interference Test (Error Recognition):
RFI Detect printed
- ☒ 5. Standard Calibration Check:
0. 100 AC
- ☒ 6. Air Blanks Completed
- ☒ 7. Timer Reset

Not a Successfully Completed Test Sequence will be printed

Instrument is operating properly and accurately. YES ☒ NO ☐

COMMENTS:

SIGNATURE K Miyazato 1139

Intoxilyzer - Alcohol Analyzer
Model 9000

SN: 90-005394

Test Date: 06/29/2025

Test Time: 04:06

Location:

Software Version: 9439.01.00

Standard Lot #: 28424100A1

Standard Value: 0.100

Expiration Date: 11/05/2026

Bottle #: 22

Last Changed By: MIYAZATO, KORY

Permit #: 30215

Operator: MIYAZATO, KORY

Permit #: 30215

Agency: GOODYEAR PD

Test	g/210L	Time
Air Blank	0.000	04:07:10
Calibration Chk	0.100	04:07:32
Air Blank	0.000	04:08:12

No RFI Detected

ARIZONA FORENSIC BREATH ALCOHOL ANALYTICAL REPORT

INSTRUMENT INFORMATION

Analytical Instrument: Intoxilyzer 9000
Serial Number: 90-005394
Software Version: 9439.01.00
Analytical Report Number: 53940629250086

QAS: MIYAZATO, KORY
QAS Permit #: 30215
Agency: GOODYEAR PD
Last 31-Day Check: 05/29/2025
Last Annual Maintenance: 08/30/2024

SUBJECT INFORMATION

Name: 31-Day Check
Test Date: 06/29/2025
Driver's License #:
Gender:
Date of Birth:
Age:

Weight:
State of Issue:
Driver's License Expiration:
Deprivation Start Time: 03:30
15 - Minute Deprivation: Yes

OPERATOR INFORMATION

Name: MIYAZATO, KORY
Agency: GOODYEAR PD

Permit #: 30215

STANDARD INFORMATION

Standard Value: 0.100
Standard Lot #: 28424100A1
Expiration Date: 11/05/2026
Bottle #: 22
Last Changed By: MIYAZATO, KORY

Permit #: 30215

Test	g/210L	Time
Air Blank	0.000	03:53:26
Diagnostic Test	PASS	03:54:02
Air Blank	0.000	03:54:39
Calibration Chk	0.100	03:55:00
Air Blank	0.000	03:55:40
Subject Test 1	DEF*	03:58:52
Air Blank	0.000	03:59:32
Air Blank	0.000	04:00:13
Subject Test 2	0.000	04:00:29
Air Blank	0.000	04:01:09
Wait		04:05:14
Air Blank	0.000	04:05:51
Subject Test 3	INV**	04:06:08
Air Blank	RFI***	04:06:21

RESULTS

EXCEPTION MESSAGES

*Deficient Sample
**Invalid Sample - Begin new deprivation period
***RFI Detect

* Deficient Sample
** Invalid Sample - Begin new deprivation period
*** RFI Detect

OPERATOR COMMENTS

Not a Successfully Completed Test Sequence

All air blank results must be 0.000.
Consecutive subject test results must not differ by more than 0.020 g/210L.
Standard check results must be $\pm 10\%$.