

EXHIBIT I-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
INTOXILYZER MODEL 9000

PERIODIC MAINTENANCE AND STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST K Miyazaki 1139 AGENCY Goodyear
DATE 6/29/25 TIME 2035
INTOXILYZER SERIAL # 90-003795

☒ 1. Ensure that gas tank is attached and contains a standard alcohol concentration 0.100 AC

DIAGNOSTIC TESTS

- ☒ 1. Clock time check
☒ 2. Date check

OPERATIONAL TESTS

- ☒ 1. Deficient Subject Test (Proper Sample Recognition)
Deficient Sample printed
- ☒ 2. Alcohol-free Subject Test (Proper Sample Recognition):
0. 000 AC
- ☒ 3. Mouth Alcohol Subject Test (Proper Sample Recognition):
Invalid Sample – Begin new deprivation period printed
- ☒ 4. Radio Frequency Interference Test (Error Recognition):
RFI Detect printed
- ☒ 5. Standard Calibration Check:
0. 101 AC
- ☒ 6. Air Blanks Completed
- ☒ 7. Timer Reset

Not a Successfully Completed Test Sequence will be printed

Instrument is operating properly and accurately. YES ☒ NO ☐

COMMENTS:

SIGNATURE K Miyazaki 1139

ARIZONA FORENSIC BREATH ALCOHOL ANALYTICAL REPORT

11 N 145TH AVE

INSTRUMENT INFORMATION

Analytical Instrument: Intoxilyzer 9000
Serial Number: 90-003795
Software Version: 9439.01.00
Analytical Report Number: 37950629250744

QAS: MIYAZATO, KORY
QAS Permit #: 30215
Agency: GOODYEAR PD
Last 31-Day Check: 06/13/2025
Last Annual Maintenance: 10/02/2024

SUBJECT INFORMATION

Name: 31-Day Check
Test Date: 06/29/2025
Driver's License #:
Gender:
Date of Birth:
Age:

Weight:
State of Issue:
Driver's License Expiration:
Deprivation Start Time: 20:15
15 - Minute Deprivation: Yes

OPERATOR INFORMATION

Name: MIYAZATO, KORY
Agency: GOODYEAR PD

Permit #: 30215

STANDARD INFORMATION

Standard Value: 0.100
Standard Lot #: 2842100A1
Expiration Date: 11/05/2026
Bottle #: 24
Last Changed By: MIYAZATO, KORY

Permit #: 30215

Test	g/210L	Time
Air Blank	0.000	20:35:09
Diagnostic Test	PASS	20:35:44
Air Blank	0.000	20:36:21
Calibration Chk	0.100	20:36:43
Air Blank	0.000	20:37:23
Subject Test 1	DEF*	20:40:34
Air Blank	0.000	20:41:14
Air Blank	0.000	20:41:55
Subject Test 2	0.000	20:42:11
Air Blank	0.000	20:42:59
Wait		20:47:04
Air Blank	0.000	20:47:41
Subject Test 3	INV**	20:48:03
Air Blank	RFI***	20:48:21

RESULTS

*Deficient Sample
**Invalid Sample - Begin new deprivation period
***RFI Detect

EXCEPTION MESSAGES

* Deficient Sample
** Invalid Sample - Begin new deprivation period
*** RFI Detect

OPERATOR COMMENTS

Not a Successfully Completed Test Sequence

*All air blank results must be 0.000.
Consecutive subject test results must not differ by more than 0.020 g/210L.
Standard check results must be $\pm 10\%$.*

Intoxilyzer - Alcohol Analyzer
Model 9000

SN: 90-003795

Test Date: 06/29/2025

Test Time: 20:48

Location: 11 N 145TH AVE

Software Version: 9439.01.00

Standard Lot #: 2842100A1

Standard Value: 0.100

Expiration Date: 11/05/2026

Bottle #: 24

Last Changed By: MIYAZATO, KORY

Permit #: 30215

Operator: MIYAZATO, KORY

Permit #: 30215

Agency: GOODYEAR PD

Test	g/210L	Time
Air Blank	0.000	20:49:13
Calibration Chk	0.101	20:49:35
Air Blank	0.000	20:50:14

No RFI Detected