

EXHIBIT I-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
INTOXILYZER MODEL 9000

PERIODIC MAINTENANCE AND STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST Michael Guilbault AGENCY GOODYEAR PD
DATE 7/29/2025 TIME 1556
INTOXILYZER SERIAL # 90-004605

- ☒ 1. Ensure that gas tank is attached and contains a standard alcohol concentration -100 AC

DIAGNOSTIC TESTS

- ☒ 1. Clock time check
☒ 2. Date check

OPERATIONAL TESTS

- ☒ 1. Deficient Subject Test (Proper Sample Recognition)
Deficient Sample printed
- ☒ 2. Alcohol-free Subject Test (Proper Sample Recognition):
0. 000 AC
- ☒ 3. Mouth Alcohol Subject Test (Proper Sample Recognition):
Invalid Sample – Begin new deprivation period printed
- ☒ 4. Radio Frequency Interference Test (Error Recognition):
RFI Detect printed
- ☐ 5. Standard Calibration Check:
0. 101 AC
- ☒ 6. Air Blanks Completed
- ☒ 7. Timer Reset

Not a Successfully Completed Test Sequence will be printed

Instrument is operating properly and accurately. YES ☒ NO ☐

COMMENTS:

SIGNATURE

MG 1144

ARIZONA FORENSIC BREATH ALCOHOL ANALYTICAL REPORT

INSTRUMENT INFORMATION

Analytical Instrument: Intoxilyzer 9000
Serial Number: 90-004605
Software Version: 9439.01.00
Analytical Report Number: 46050729250572

QAS: MIYAZATO, KORY
QAS Permit #: 30215
Agency: GOODYEAR PD
Last 31-Day Check: 07/05/2025
Last Annual Maintenance: 06/16/2025

SUBJECT INFORMATION

Name: 31-Day Check
Test Date: 07/29/2025
Driver's License #:
Gender:
Date of Birth:
Age:

Weight:
State of Issue:
Driver's License Expiration:
Deprivation Start Time: 15:40
15 - Minute Deprivation: Yes

OPERATOR INFORMATION

Name: GUILBAULT, MICHAEL
Agency: GOODYEAR PD

Permit #: 30305

STANDARD INFORMATION

Standard Value: 0.100
Standard Lot #: 28424100A1
Expiration Date: 11/05/2026
Bottle #: 024
Last Changed By: MIYAZATO, KORY

Permit #: 30215

Test	g/210L	Time
Air Blank	0.000	15:56:02
Diagnostic Test	PASS	15:56:38
Air Blank	0.000	15:57:15
Calibration Chk	0.100	15:57:37
Air Blank	0.000	15:58:16
Subject Test 1	DEF*	16:01:29
Air Blank	0.000	16:02:21
Air Blank	0.000	16:03:01
Subject Test 2	0.000	16:03:28
Air Blank	0.000	16:04:17
Wait		16:08:20
Air Blank	0.000	16:08:58
Subject Test 3	INV**	16:09:47
Air Blank	RFI***	16:09:58

*Deficient Sample
**Invalid Sample - Begin new deprivation period
***RFI Detect

RESULTS

EXCEPTION MESSAGES

* Deficient Sample
** Invalid Sample - Begin new deprivation period
*** RFI Detect

OPERATOR COMMENTS

31` DAY CHECK QAS

Not a Successfully Completed Test Sequence

All air blank results must be 0.000.
Consecutive subject test results must not differ by more than 0.020 g/210L.
Standard check results must be $\pm 10\%$.

Intoxilyzer - Alcohol Analyzer
Model 9000

SN: 90-004605

Test Date: 07/29/2025

Test Time: 16:10

Location:

Software Version: 9439.01.00

Standard Lot #: 28424100A1

Standard Value: 0.100

Expiration Date: 11/05/2026

Bottle #: 024

Last Changed By: MIYAZATO, KORY

Permit #: 30215

Operator: GUILBAULT, MICHAEL

Permit #: 30305

Agency: GOODYEAR PD

Test	g/210L	Time
Air Blank	0.000	16:11:03
Calibration Chk	0.101	16:11:26
Air Blank	0.000	16:12:05

No RFI Detected