

EXHIBIT I-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
INTOXILYZER MODEL 9000

PERIODIC MAINTENANCE AND STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST K Miyazawa 1139 AGENCY Goodyear
DATE 9/9/25 TIME 2058
INTOXILYZER SERIAL # 90-003795

- ☒ 1. Ensure that gas tank is attached and contains a standard alcohol concentration 0.100 AC

DIAGNOSTIC TESTS

- ☒ 1. Clock time check
☒ 2. Date check

OPERATIONAL TESTS

- ☒ 1. Deficient Subject Test (Proper Sample Recognition)
Deficient Sample printed
☒ 2. Alcohol-free Subject Test (Proper Sample Recognition):
0. 000 AC
☒ 3. Mouth Alcohol Subject Test (Proper Sample Recognition):
Invalid Sample – Begin new deprivation period printed
☒ 4. Radio Frequency Interference Test (Error Recognition):
RFI Detect printed
☒ 5. Standard Calibration Check:
0. 093 AC
☒ 6. Air Blanks Completed
☒ 7. Timer Reset

Not a Successfully Completed Test Sequence will be printed

Instrument is operating properly and accurately. YES ☒ NO ☐

COMMENTS:

SIGNATURE Kurt 1139

ARIZONA FORENSIC BREATH ALCOHOL ANALYTICAL REPORT

10701 S ESTRELLA PWY

INSTRUMENT INFORMATION

Analytical Instrument: Intoxilyzer 9000
Serial Number: 90-003795
Software Version: 9439.01.00
Analytical Report Number: 37950909250765

QAS: MIYAZATO, KORY
QAS Permit #: 30215
Agency: GOODYEAR PD
Last 31-Day Check: 08/12/2025
Last Annual Maintenance: 10/02/2024

SUBJECT INFORMATION

Name: 31-Day Check
Test Date: 09/09/2025
Driver's License #:
Gender:
Date of Birth:
Age:

Weight:
State of Issue:
Driver's License Expiration:
Deprivation Start Time: 20:30
15 - Minute Deprivation: Yes

OPERATOR INFORMATION

Name: MIYAZATO, KORY
Agency: GOODYEAR PD

Permit #: 30215

STANDARD INFORMATION

Standard Value: 0.100
Standard Lot #: 302402972207
Expiration Date: 02/20/2026
Bottle #: 19
Last Changed By: MIYAZATO, KORY

Permit #: 30215

Test	g/210L	Time
Air Blank	0.000	20:58:46
Diagnostic Test	PASS	20:59:23
Air Blank	0.000	21:00:00
Calibration Chk	0.094	21:00:22
Air Blank	0.000	21:01:01
Subject Test 1	DEF*	21:04:18
Air Blank	0.000	21:05:06
Air Blank	0.000	21:05:46
Subject Test 2	0.000	21:06:02
Air Blank	0.000	21:06:45
Wait		21:10:49
Air Blank	0.000	21:11:26
Subject Test 3	INV**	21:11:44
Air Blank	RFI***	21:11:57

*Deficient Sample
**Invalid Sample - Begin new deprivation period
***RFI Detect

RESULTS

EXCEPTION MESSAGES

* Deficient Sample
** Invalid Sample - Begin new deprivation period
*** RFI Detect

OPERATOR COMMENTS

Not a Successfully Completed Test Sequence

All air blank results must be 0.000.
Consecutive subject test results must not differ by more than 0.020 g/210L.
Standard check results must be $\pm 10\%$.

Intoxilyzer - Alcohol Analyzer
Model 9000

SN: 90-003795
Test Date: 09/09/2025
Location: 10701 S ESTRELLA PWY
Test Time: 21:12

Software Version: 9439.01.00

Standard Lot #: 302402972207
Standard Value: 0.100
Expiration Date: 02/20/2026
Bottle #: 19
Last Changed By: MIYAZATO, KORY
Permit #: 30215

Operator: MIYAZATO, KORY
Permit #: 30215
Agency: GOODYEAR PD

Test	g/210L	Time
Air Blank	0.000	21:12:51
Calibration Chk	0.093	21:13:13
Air Blank	0.000	21:13:53

No RFI Detected