

EXHIBIT I-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
INTOXILYZER MODEL 9000

PERIODIC MAINTENANCE AND STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST K Miyazato ¹¹³⁹ AGENCY Goodyear
DATE 10/30/25 TIME 0428
INTOXILYZER SERIAL # 90-005394

☒ 1. Ensure that gas tank is attached and contains a standard alcohol concentration 100 AC

DIAGNOSTIC TESTS

- ☒ 1. Clock time check
☒ 2. Date check

OPERATIONAL TESTS

☒ 1. Deficient Subject Test (Proper Sample Recognition)

Deficient Sample printed

☒ 2. Alcohol-free Subject Test (Proper Sample Recognition):

0. 000 AC

☒ 3. Mouth Alcohol Subject Test (Proper Sample Recognition):

Invalid Sample – Begin new deprivation period printed

☒ 4. Radio Frequency Interference Test (Error Recognition):

RFI Detect printed

☒ 5. Standard Calibration Check:

0. 100 AC

☒ 6. Air Blanks Completed

☒ 7. Timer Reset

Not a Successfully Completed Test Sequence will be printed

Instrument is operating properly and accurately. YES ☒ NO ☐

COMMENTS:

SIGNATURE KMT ¹¹³⁹

ARIZONA FORENSIC BREATH ALCOHOL ANALYTICAL REPORT

INSTRUMENT INFORMATION

Analytical Instrument: Intoxilyzer 9000
Serial Number: 90-005394
Software Version: 9439.01.00
Analytical Report Number: 53941030250118

QAS: MIYAZATO, KORY
QAS Permit #: 30215
Agency: GOODYEAR PD
Last 31-Day Check: 10/01/2025
Last Annual Maintenance: 09/04/2025

SUBJECT INFORMATION

Name: 31-Day Check
Test Date: 10/30/2025
Driver's License #:
Gender:
Date of Birth:
Age:

Weight:
State of Issue:
Driver's License Expiration:
Deprivation Start Time: 04:10
15 - Minute Deprivation: Yes

OPERATOR INFORMATION

Name: MIYAZATO, KORY
Agency: GOODYEAR PD

Permit #: 30215

STANDARD INFORMATION

Standard Value: 0.100
Standard Lot #: 28424100A1
Expiration Date: 11/05/2026
Bottle #: 22
Last Changed By: MIYAZATO, KORY

Permit #: 30215

Test	g/210L	Time
Air Blank	0.000	04:28:13
Diagnostic Test	PASS	04:28:49
Air Blank	0.000	04:29:26
Calibration Chk	0.100	04:29:47
Air Blank	0.000	04:30:27
Subject Test 1	DEF*	04:33:44
Air Blank	0.000	04:34:37
Air Blank	0.000	04:35:17
Subject Test 2	0.000	04:35:33
Air Blank	0.000	04:36:14
Wait		04:40:18
Air Blank	0.000	04:40:55
Subject Test 3	INV**	04:41:29
Air Blank	RFI***	04:41:45

RESULTS

EXCEPTION MESSAGES

*Deficient Sample
**Invalid Sample - Begin new deprivation period
***RFI Detect

* Deficient Sample
** Invalid Sample - Begin new deprivation period
*** RFI Detect

OPERATOR COMMENTS

Not a Successfully Completed Test Sequence

All air blank results must be 0.000.
Consecutive subject test results must not differ by more than 0.020 g/210L.
Standard check results must be $\pm 10\%$.

Intoxilyzer - Alcohol Analyzer
Model 9000

SN: 90-005394

Test Date: 10/30/2025

Test Time: 04:42

Location:

Software Version: 9439.01.00

Standard Lot #: 28424100A1

Standard Value: 0.100

Expiration Date: 11/05/2026

Bottle #: 22

Last Changed By: MIYAZATO, KORY

Permit #: 30215

Operator: MIYAZATO, KORY

Permit #: 30215

Agency: GOODYEAR PD

Test	g/210L	Time
Air Blank	0.000	04:42:34
Calibration Chk	0.100	04:42:56
Air Blank	0.000	04:43:35

No RFI Detected