

EXHIBIT I-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
INTOXILYZER MODEL 9000

PERIODIC MAINTENANCE AND STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST K M 1139 112410 AGENCY Goodyear
DATE 10/30/25 TIME 0511
INTOXILYZER SERIAL # 90-004605

☒ 1. Ensure that gas tank is attached and contains a standard alcohol concentration 100 AC

DIAGNOSTIC TESTS

- ☒ 1. Clock time check
☒ 2. Date check

OPERATIONAL TESTS

- ☒ 1. Deficient Subject Test (Proper Sample Recognition)

Deficient Sample printed

- ☒ 2. Alcohol-free Subject Test (Proper Sample Recognition):

0. 000 AC

- ☒ 3. Mouth Alcohol Subject Test (Proper Sample Recognition):

Invalid Sample – Begin new deprivation period printed

- ☒ 4. Radio Frequency Interference Test (Error Recognition):

RFI Detect printed

- ☒ 5. Standard Calibration Check:

0. 100 AC

- ☒ 6. Air Blanks Completed

- ☒ 7. Timer Reset

Not a Successfully Completed Test Sequence will be printed

Instrument is operating properly and accurately.

YES ☒ NO ☐

COMMENTS:

SIGNATURE

K M 1139

ARIZONA FORENSIC BREATH ALCOHOL ANALYTICAL REPORT

INSTRUMENT INFORMATION

Analytical Instrument: Intoxilyzer 9000
Serial Number: 90-004605
Software Version: 9439.01.00
Analytical Report Number: 46051030250709

QAS: MIYAZATO, KORY
QAS Permit #: 30215
Agency: GOODYEAR PD
Last 31-Day Check: 10/01/2025
Last Annual Maintenance: 06/16/2025

SUBJECT INFORMATION

Name: 31-Day Check
Test Date: 10/30/2025
Driver's License #:
Gender:
Date of Birth:
Age:

Weight:
State of Issue:
Driver's License Expiration:
Deprivation Start Time: 04:45
15 - Minute Deprivation: Yes

OPERATOR INFORMATION

Name: MIYAZATO, KORY
Agency: GOODYEAR PD

Permit #: 30215

STANDARD INFORMATION

Standard Value: 0.100
Standard Lot #: 28424100A1
Expiration Date: 11/05/2026
Bottle #: 023
Last Changed By: MIYAZATO, KORY

Permit #: 30215

Test	g/210L	Time
Air Blank	0.000	05:11:33
Diagnostic Test	PASS	05:12:09
Air Blank	0.000	05:12:46
Calibration Chk	0.101	05:13:08
Air Blank	0.000	05:13:47
Subject Test 1	DEF*	05:16:58
Air Blank	0.000	05:17:42
Air Blank	0.000	05:18:22
Subject Test 2	INV**	05:18:39
Air Blank	RFI***	05:18:52

RESULTS

*Deficient Sample
**Invalid Sample - Begin new deprivation period
***RFI Detect

EXCEPTION MESSAGES

* Deficient Sample
** Invalid Sample - Begin new deprivation period
*** RFI Detect

OPERATOR COMMENTS

Not a Successfully Completed Test Sequence

*All air blank results must be 0.000.
Consecutive subject test results must not differ by more than 0.020 g/210L.
Standard check results must be $\pm 10\%$.*

Intoxilyzer - Alcohol Analyzer

Model 9000

SN: 90-004605

Test Date: 10/30/2025

Test Time: 05:19

Location:

Software Version: 9439.01.00

Standard Lot #: 28424100A1

Standard Value: 0.100

Expiration Date: 11/05/2026

Bottle #: 023

Last Changed By: MIYAZATO, KORY

Permit #: 30215

Operator: MIYAZATO, KORY

Permit #: 30215

Agency: GOODYEAR PD

<u>Test</u>	<u>g/210L</u>	<u>Time</u>
Air Blank	0.000	05:19:39
Calibration Chk	0.100	05:20:01
Air Blank	0.000	05:20:41

No RFI Detected