

**EXHIBIT I-2**  
**THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)**

**ARIZONA DEPARTMENT OF PUBLIC SAFETY**  
**INTOXILYZER MODEL 9000**

**PERIODIC MAINTENANCE AND STANDARD QUALITY ASSURANCE PROCEDURE**

QA SPECIALIST K Miyazawa 1138 AGENCY Goodyear  
DATE 12/2/2005 TIME 2149  
INTOXILYZER SERIAL # 90-004605

☒ 1. Ensure that gas tank is attached and contains a standard alcohol concentration .100 AC

**DIAGNOSTIC TESTS**

- ☒ 1. Clock time check  
☒ 2. Date check

**OPERATIONAL TESTS**

- ☒ 1. Deficient Subject Test (Proper Sample Recognition)  
Deficient Sample printed
- ☒ 2. Alcohol-free Subject Test (Proper Sample Recognition):  
0. 0.00 AC
- ☒ 3. Mouth Alcohol Subject Test (Proper Sample Recognition):  
Invalid Sample – Begin new deprivation period printed
- ☒ 4. Radio Frequency Interference Test (Error Recognition):  
RFI Detect printed
- ☒ 5. Standard Calibration Check:  
0. 0.00 AC
- ☒ 6. Air Blanks Completed
- ☒ 7. Timer Reset

Not a Successfully Completed Test Sequence will be printed

Instrument is operating properly and accurately. YES ☒ NO ☐

**COMMENTS:**

CONDUCTED ANOTHER QAS DUE TO THANKSGIVING  
HOLIDAY AND EARLY MONTHLY CHECK

SIGNATURE K Miyazawa

# ARIZONA FORENSIC BREATH ALCOHOL ANALYTICAL REPORT

## INSTRUMENT INFORMATION

Analytical Instrument: Intoxilyzer 9000  
Serial Number: 90-004605  
Software Version: 9439.01.00  
Analytical Report Number: 46051202250765

QAS: MIYAZATO, KORY  
QAS Permit #: 30215  
Agency: GOODYEAR PD  
Last 31-Day Check: 11/26/2025  
Last Annual Maintenance: 06/16/2025

## SUBJECT INFORMATION

Name: 31-Day Check  
Test Date: 12/02/2025  
Driver's License #:  
Gender:  
Date of Birth:  
Age:

Weight:  
State of Issue:  
Driver's License Expiration:  
Deprivation Start Time: 21:30  
15 - Minute Deprivation: Yes

## OPERATOR INFORMATION

Name: MIYAZATO, KORY  
Agency: GOODYEAR PD

Permit #: 30215  
STANDARD INFORMATION  
Standard Value: 0.100  
Standard Lot #: 28424100A1  
Expiration Date: 11/05/2026  
Bottle #: 023  
Last Changed By: MIYAZATO, KORY

Permit #: 30215

| Test            | g/210L | Time     |
|-----------------|--------|----------|
| Air Blank       | 0.000  | 21:49:06 |
| Diagnostic Test | PASS   | 21:49:42 |
| Air Blank       | 0.000  | 21:50:19 |
| Calibration Chk | 0.101  | 21:50:41 |
| Air Blank       | 0.000  | 21:51:21 |
| Subject Test 1  | DEF*   | 21:54:45 |
| Air Blank       | 0.000  | 21:55:26 |
| Air Blank       | 0.000  | 21:56:06 |
| Subject Test 2  | 0.000  | 21:56:21 |
| Air Blank       | 0.000  | 21:57:01 |
| Wait            |        | 22:01:05 |
| Air Blank       | 0.000  | 22:01:43 |
| Subject Test 3  | INV**  | 22:02:04 |
| Air Blank       | RFI*** | 22:02:17 |

\*Deficient Sample  
\*\*Invalid Sample - Begin new deprivation period  
\*\*\*RFI Detect

## RESULTS

## EXCEPTION MESSAGES

\* Deficient Sample  
\*\* Invalid Sample - Begin new deprivation period  
\*\*\* RFI Detect

## OPERATOR COMMENTS

Not a Successfully Completed Test Sequence

*All air blank results must be 0.000.  
Consecutive subject test results must not differ by more than 0.020 g/210L.  
Standard check results must be  $\pm 10\%$ .*

Intoxilyzer - Alcohol Analyzer

Model 9000

SN: 90-004605

Test Date: 12/02/2025

Test Time: 22:02

Location:

Software Version: 9439.01.00

Standard Lot #: 28424100A1

Standard Value: 0.100

Expiration Date: 11/05/2026

Bottle #: 023

Last Changed By: MIYAZATO, KORY

Permit #: 30215

Operator: MIYAZATO, KORY

Permit #: 30215

Agency: GOODYEAR PD

| Test            | g/210L | Time     |
|-----------------|--------|----------|
| Air Blank       | 0.000  | 22:03:05 |
| Calibration Chk | 0.100  | 22:03:26 |
| Air Blank       | 0.000  | 22:04:06 |

No RFI Detected