

EXHIBIT I-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
INTOXILYZER MODEL 9000

PERIODIC MAINTENANCE AND STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST K Miyazato 1139 AGENCY Goodyear PD
DATE 12/2/2025 TIME 2143
INTOXILYZER SERIAL # 90-005394

- ☒ 1. Ensure that gas tank is attached and contains a standard alcohol concentration .100 AC

DIAGNOSTIC TESTS

- ☒ 1. Clock time check
☒ 2. Date check

OPERATIONAL TESTS

- ☒ 1. Deficient Subject Test (Proper Sample Recognition)
Deficient Sample printed
☒ 2. Alcohol-free Subject Test (Proper Sample Recognition):
0. 000 AC
☒ 3. Mouth Alcohol Subject Test (Proper Sample Recognition):
Invalid Sample – Begin new deprivation period printed
☒ 4. Radio Frequency Interference Test (Error Recognition):
RFI Detect printed
☒ 5. Standard Calibration Check:
0. 100 AC
☒ 6. Air Blanks Completed
☒ 7. Timer Reset

Not a Successfully Completed Test Sequence will be printed

Instrument is operating properly and accurately. YES ☒ NO ☐

COMMENTS:

SIGNATURE K Miyazato 1139

ARIZONA FORENSIC BREATH ALCOHOL ANALYTICAL REPORT

INSTRUMENT INFORMATION

Analytical Instrument: Intoxilyzer 9000
Serial Number: 90-005394
Software Version: 9439.01.00
Analytical Report Number: 53941202250123

QAS: MIYAZATO, KORY
QAS Permit #: 30215
Agency: GOODYEAR PD
Last 31-Day Check: 10/30/2025
Last Annual Maintenance: 09/04/2025

SUBJECT INFORMATION

Name: 31-Day Check
Test Date: 12/02/2025
Driver's License #:
Gender:
Date of Birth:
Age:

Weight:
State of Issue:
Driver's License Expiration:
Deprivation Start Time: 21:20
15 - Minute Deprivation: Yes

OPERATOR INFORMATION

Name: MIYAZATO, KORY
Agency: GOODYEAR PD

Permit #: 30215

STANDARD INFORMATION

Standard Value: 0.100
Standard Lot #: 28424100A1
Expiration Date: 11/05/2026
Bottle #: 22
Last Changed By: MIYAZATO, KORY

Permit #: 30215

Test	g/210L	Time
Air Blank	0.000	21:43:10
Diagnostic Test	PASS	21:43:45
Air Blank	0.000	21:44:22
Calibration Chk	0.102	21:44:44
Air Blank	0.000	21:45:24
Subject Test 1	DEF*	21:48:35
Air Blank	0.000	21:49:17
Air Blank	0.000	21:49:57
Subject Test 2	0.000	21:50:12
Air Blank	0.000	21:50:53
Wait		21:54:56
Air Blank	0.000	21:55:33
Subject Test 3	INV**	21:55:53
Air Blank	RFI***	21:56:09

RESULTS

EXCEPTION MESSAGES

*Deficient Sample
**Invalid Sample - Begin new deprivation period
***RFI Detect

* Deficient Sample
** Invalid Sample - Begin new deprivation period
*** RFI Detect

OPERATOR COMMENTS

Not a Successfully Completed Test Sequence

*All air blank results must be 0.000.
Consecutive subject test results must not differ by more than 0.020 g/210L.
Standard check results must be $\pm 10\%$.*

Intoxilyzer - Alcohol Analyzer

Model 9000

SN: 90-005394

Test Date: 12/02/2025

Test Time: 21:56

Location:

Software Version: 9439.01.00

Standard Lot #: 28424100A1

Standard Value: 0.100

Expiration Date: 11/05/2026

Bottle #: 22

Last Changed By: MIYAZATO, KORY

Permit #: 30215

Operator: MIYAZATO, KORY

Permit #: 30215

Agency: GOODYEAR PD

Test	g/210L	Time
Air Blank	0.000	21:57:01
Calibration Chk	0.100	21:57:23
Air Blank	0.000	21:58:02

No RFI Detected