

EXHIBIT I-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)
ARIZONA DEPARTMENT OF PUBLIC SAFETY
INTOXILYZER MODEL 9000

PERIODIC MAINTENANCE AND STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST E. JOHNSON #1187 AGENCY GOODYEAR PD
DATE 05-31-2026 TIME 2300
INTOXILYZER SERIAL # 90-004605

☒ 1. Ensure that gas tank is attached and contains a standard alcohol concentration 0.100 AC

DIAGNOSTIC TESTS

- ☒ 1. Clock time check
- ☒ 2. Date check

OPERATIONAL TESTS

- ☒ 1. Deficient Subject Test (Proper Sample Recognition)
Deficient Sample printed
- ☒ 2. Alcohol-free Subject Test (Proper Sample Recognition):
0.000 AC
- ☒ 3. Mouth Alcohol Subject Test (Proper Sample Recognition):
Invalid Sample – Begin new deprivation period printed
- ☒ 4. Radio Frequency Interference Test (Error Recognition):
RFI Detect printed
- ☒ 5. Standard Calibration Check:
0.099 AC
- ☒ 6. Air Blanks Completed
- ☒ 7. Timer Reset

Not a Successfully Completed Test Sequence will be printed

Instrument is operating properly and accurately. YES ☒ NO ☐

COMMENTS:

SIGNATURE ERIK JOHNSON Digitally signed by ERIK JOHNSON
Date: 2026.05.31 23:32:30 -07'00'

ARIZONA FORENSIC BREATH ALCOHOL ANALYTICAL REPORT

INSTRUMENT INFORMATION

Analytical Instrument: Intoxilyzer 9000
Serial Number: 90-004605
Software Version: 9439.01.00
Analytical Report Number: 46050531260888

QAS: JOHNSON, ERIK
QAS Permit #: 30455
Agency: GOODYEAR PD
Last 31-Day Check: 05/31/2026
Last Annual Maintenance: 06/16/2025

SUBJECT INFORMATION

Name: 31-Day Check
Test Date: 05/31/2026
Driver's License #:
Gender:
Date of Birth:
Age:

Weight:
State of Issue:
Driver's License Expiration:
Deprivation Start Time: 22:00
15 - Minute Deprivation: Yes

OPERATOR INFORMATION

Name: JOHNSON, ERIK
Agency: GOODYEAR PD

Permit #: 30455

STANDARD INFORMATION

Standard Value: 0.100
Standard Lot #: 28424100A1
Expiration Date: 11/05/2026
Bottle #: 025
Last Changed By: MIYAZATO, KORY

Permit #: 30215

Test	g/210L	Time
Air Blank	0.000	23:00:31
Diagnostic Test	PASS	23:01:07
Air Blank	0.000	23:01:44
Calibration Chk	0.099	23:02:05
Air Blank	0.000	23:02:45
Subject Test 1	DEF*	23:05:57
Air Blank	0.000	23:06:45
Air Blank	0.000	23:07:25
Subject Test 2	0.000	23:07:52
Air Blank	0.000	23:08:39
Wait		23:12:42
Air Blank	0.000	23:13:20
Subject Test 3	INV**	23:13:45
Air Blank	RFI***	23:13:58

RESULTS

EXCEPTION MESSAGES

*Deficient Sample
**Invalid Sample - Begin new deprivation period
***RFI Detect

* Deficient Sample
** Invalid Sample - Begin new deprivation period
*** RFI Detect

OPERATOR COMMENTS

Not a Successfully Completed Test Sequence

All air blank results must be 0.000.
Consecutive subject test results must not differ by more than 0.020 g/210L.
Standard check results must be $\pm 10\%$.