

Intoxilyzer - Alcohol Analyzer
Model 9000

SN: 90-003795

Test Date: 05/25/2026

Test Time: 20:44

Location: 10701 S ESTRELLA PWY

Software Version: 9439.01.00

Standard Lot #: 21125100A1

Standard Value: 0.100

Expiration Date: 09/05/2027

Bottle #: 025

Last Changed By: JOHNSON, ERIK

Permit #: 30455

Operator: JOHNSON, ERIK

Permit #: 30455

Agency: GOODYEAR PD

Test	g/210L	Time
Air Blank	0.000	20:44:46
Calibration Chk	0.100	20:45:08
Air Blank	0.000	20:45:48

No RFI Detected

Successfully Completed Test Sequence



7 Eastgate Dr. • P.O. Box 790 • Jacksonville, IL 62651-0790
217-245-2183 • Fax: 217-243-7634 • www.ilmoproducts.com

Certificate of Analysis

Certificate ID: 17454
Part #: BAC105L100T
Cylinder Size: 105L
Lot Number: 21125100A1
Expiration: 9/5/2027

0.100 BAC (For the calibration of instruments used to determine breath alcohol concentration)

Contents: 105 Liters @ 1000 psig 70°F (21°C)


Component:	Reported Concentration:	Analytical Accuracy (U, k=2): abs	Analytical Method:
Ethanol	260 ppm	+/-0.002 BAC(G/210L) [5.2 ppm]	NDIR
Nitrogen	Balance		

Distributed by:


CMI Inc.
316 East Ninth Street
Owensboro, KY 42303
Phone 866-835-0690
www.alcoholtest.com

*Reference Standard:
NIST Traceable Certified Reference Material - 260.6 µmol/mol
Ethanol in Nitrogen - Serial No. GN0015020 Lot No. 00424E1

Store in dry area, away from sources of heat,
ignition and direct sunlight. Do not allow storage
area to exceed 52 °C (125 °F).


Specialty Gas Lab Tech

08-14-2025
Issuance Date



The calibration results within this certificate were obtained at the facility listed above using equipment and standards capable of producing analytical results traceable to NIST, and apply only to the items contained on this certificate. ILMO Products Company makes no warranty or representation as to the suitability of the use of any information provided for any particular purpose. The information use is at the sole discretion and risk of the user. Liability shall be limited to established replacement cost of this material or service. This certificate applies only to the items described and shall not be reproduced other than in full, without written approval from the issuing facility.

ISO/IEC 17025:2017 Accredited Laboratory