

EXHIBIT I-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
INTOXILYZER MODEL 9000

PERIODIC MAINTENANCE AND STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST E. JOHNSON #1187 AGENCY GOODYEAR PD
DATE 06-30-2026 TIME 1925
INTOXILYZER SERIAL # 90-003795

- ☒ 1. Ensure that gas tank is attached and contains a standard alcohol concentration 0.100 AC

DIAGNOSTIC TESTS

- ☒ 1. Clock time check
☒ 2. Date check

OPERATIONAL TESTS

- ☒ 1. Deficient Subject Test (Proper Sample Recognition)
Deficient Sample printed
- ☒ 2. Alcohol-free Subject Test (Proper Sample Recognition):
0.000 AC
- ☒ 3. Mouth Alcohol Subject Test (Proper Sample Recognition):
Invalid Sample – Begin new deprivation period printed
- ☒ 4. Radio Frequency Interference Test (Error Recognition):
RFI Detect printed
- ☒ 5. Standard Calibration Check:
0.099 AC
- ☒ 6. Air Blanks Completed
- ☒ 7. Timer Reset

Not a Successfully Completed Test Sequence will be printed

Instrument is operating properly and accurately. YES ☒ NO ☐

COMMENTS:

SIGNATURE ERIK JOHNSON Digitally signed by ERIK JOHNSON
Date: 2026.06.30 19:44:11 -07'00'

ARIZONA FORENSIC BREATH ALCOHOL ANALYTICAL REPORT

11 N 145TH AVE

INSTRUMENT INFORMATION

Analytical Instrument: Intoxilyzer 9000
Serial Number: 90-003795
Software Version: 9439.01.00
Analytical Report Number: 37950630260808

QAS: JOHNSON, ERIK
QAS Permit #: 30455
Agency: GOODYEAR PD
Last 31-Day Check: 05/31/2026
Last Annual Maintenance: 10/03/2025

SUBJECT INFORMATION

Name: 31-Day Check
Test Date: 06/30/2026
Driver's License #:
Gender:
Date of Birth:
Age:

Weight:
State of Issue:
Driver's License Expiration:
Deprivation Start Time: 18:00
15 - Minute Deprivation: Yes

OPERATOR INFORMATION

Name: JOHNSON, ERIK
Agency: GOODYEAR PD

Permit #: 30455

STANDARD INFORMATION

Standard Value: 0.100
Standard Lot #: 21125100A1
Expiration Date: 09/05/2027
Bottle #: 025
Last Changed By: JOHNSON, ERIK

Permit #: 30455

Test	g/210L	Time
Air Blank	0.000	19:25:34
Diagnostic Test	PASS	19:26:09
Air Blank	0.000	19:26:46
Calibration Chk	0.100	19:27:08
Air Blank	0.000	19:27:48
Subject Test 1	DEF*	19:31:06
Air Blank	0.000	19:32:06
Air Blank	0.000	19:32:46
Subject Test 2	0.000	19:33:09
Air Blank	0.000	19:33:51
Wait		19:37:54
Air Blank	0.000	19:38:31
Subject Test 3	INV**	19:39:21
Air Blank	RFI***	19:39:34

*Deficient Sample
**Invalid Sample - Begin new deprivation period
***RFI Detect

RESULTS

EXCEPTION MESSAGES

* Deficient Sample
** Invalid Sample - Begin new deprivation period
*** RFI Detect

OPERATOR COMMENTS

Not a Successfully Completed Test Sequence

All air blank results must be 0.000.
Consecutive subject test results must not differ by more than 0.020 g/210L.
Standard check results must be $\pm 10\%$.

Intoxilyzer - Alcohol Analyzer
Model 9000

SN: 90-003795

Test Date: 06/30/2026

Test Time: 19:40

Location: 11 N 145TH AVE

Software Version: 9439.01.00

Standard Lot #: 21125100A1

Standard Value: 0.100

Expiration Date: 09/05/2027

Bottle #: 025

Last Changed By: JOHNSON, ERIK

Permit #: 30455

Operator: JOHNSON, ERIK

Permit #: 30455

Agency: GOODYEAR PD

Test	g/210L	Time
Air Blank	0.000	19:40:53
Calibration Chk	0.099	19:41:14
Air Blank	0.000	19:41:54

No RFI Detected