

EXHIBIT I-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
INTOXILYZER MODEL 9000

PERIODIC MAINTENANCE AND STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST Jason Howard #1489 AGENCY GYPD
DATE 07/31/2024 TIME 0431
INTOXILYZER SERIAL # 90-004605

- ☒ 1. Ensure that gas tank is attached and contains a standard alcohol concentration 0.100 AC

DIAGNOSTIC TESTS

- ☐ 1. Clock time check
☒ 2. Date check

OPERATIONAL TESTS

- ☒ 1. Deficient Subject Test (Proper Sample Recognition)
Deficient Sample printed
☒ 2. Alcohol-free Subject Test (Proper Sample Recognition):
0. 000 AC
☒ 3. Mouth Alcohol Subject Test (Proper Sample Recognition):
Invalid Sample – Begin new deprivation period printed
☒ 4. Radio Frequency Interference Test (Error Recognition):
RFI Detect printed
☒ 5. Standard Calibration Check:
0. 098 AC
☒ 6. Air Blanks Completed
☒ 7. Timer Reset

Not a Successfully Completed Test Sequence will be printed

Instrument is operating properly and accurately. YES ☒ NO ☐

COMMENTS:

SIGNATURE J. Howard #1489

ARIZONA FORENSIC BREATH ALCOHOL ANALYTICAL REPORT

INSTRUMENT INFORMATION

Analytical Instrument: Intoxilyzer 9000
Serial Number: 90-004605
Software Version: 9439.01.00
Analytical Report Number: 46050731260912

QAS: JOHNSON, ERIK
QAS Permit #: 30455
Agency: GOODYEAR PD
Last 31-Day Check: 06/30/2026
Last Annual Maintenance: 06/17/2026

SUBJECT INFORMATION

Name: 31-Day Check
Test Date: 07/31/2026
Driver's License #:
Gender:
Date of Birth:
Age:

Weight:
State of Issue:
Driver's License Expiration:
Deprivation Start Time: 04:00
15 - Minute Deprivation: Yes

OPERATOR INFORMATION

Name: HOWARD, JASON
Agency: GOODYEAR PD

Permit #: 30447

STANDARD INFORMATION

Standard Value: 0.100
Standard Lot #: 28424100A1
Expiration Date: 11/05/2026
Bottle #: 025
Last Changed By: MIYAZATO, KORY

Permit #: 30215

Test	g/210L	Time
Air Blank	0.000	04:45:43
Diagnostic Test	PASS	04:46:19
Air Blank	0.000	04:46:56
Calibration Chk	0.098	04:47:17
Air Blank	0.000	04:47:57
Subject Test 1	DEF*	04:51:45
Air Blank	0.000	04:52:32
Air Blank	0.000	04:53:12
Subject Test 2	0.000	04:53:44
Air Blank	0.000	04:54:27
Wait		04:58:31
Air Blank	0.000	04:59:08
Subject Test 3	INV**	05:04:09
Air Blank	RFI***	05:04:22

*Deficient Sample
**Invalid Sample - Begin new deprivation period
***RFI Detect

RESULTS

EXCEPTION MESSAGES

* Deficient Sample
** Invalid Sample - Begin new deprivation period
*** RFI Detect

OPERATOR COMMENTS

Not a Successfully Completed Test Sequence

All air blank results must be 0.000.
Consecutive subject test results must not differ by more than 0.020 g/210L.
Standard check results must be $\pm 10\%$.